

Statement on the use of the prophylaxis powders AIR POLISHING POWDER BICARBONATE, AIR POLISHING POWDER GLYCINE, AIR POLISHING POWDER TAGATOSE with market-proven air polishing devices.

Statement zur Verwendung der Prophylaxepulver AIR POLISHING POWDER BICARBONATE, AIR POLISHING POWDER GLYCINE, AIR POLISHING POWDER TAGATOSE mit am Markt gängigen Pulverstrahlgeräten.

The prophylaxis powders AIR POLISHING POWDER BICARBONATE, AIR POLISHING POWDER GLYCINE, AIR POLISHING POWDER TAGATOSE have been tested with the following market-proven air polishing devices: Airflow Handy 2+, Airflow Handy 3.0 Perio Premium, Airflow One, Airflow Master (EMS Electro Medical Systems SA, Switzerland); Prophy-Mate neo¹, Perio-Mate² (NSK, Nakanishi Inc., Japan); Combi Touch, Starjet³ (Mectron s.p.a., Italy); Air Max (Acteon Germany GmbH, Germany); PROPHY Flex 3 (KaVo Dental GmbH, Germany). The manufacturer's instructions manual of the device were followed. The powders listed above are compatible and approved with the tested air polishing devices. Since the design and functionality of other market-proven devices is comparable to those tested, it can be assumed that our prophylaxis powders can be used without problems in all devices and that the **release** therefore extends to all devices of the same or similar design.

In general, the instruction manual of the device manufacturer as well as the instruction manual of the prophylaxis powder must be observed.

Die Prophylaxepulver AIR POLISHING POWDER BICARBONATE, AIR POLISHING POWDER GLYCINE, AIR POLISHING POWDER TAGATOSE wurden mit den folgenden am Markt gängigen Pulverstrahlgeräten:

¹ The device is identical in construction to the Smartgun One (Omnident Dental-Handelsgesellschaft mbH, Germany) device.

² The device is identical in construction to the Smartgun Perio (Omnident Dental-Handelsgesellschaft mbH, Germany) device.

³ The device is identical in construction to the Proxeo Aura (W&H Dentalwerk Bürmoos GmbH, Austria) device.

PRISMAN GmbH • Otto-Hahn-Ring 6-18 • D-64653 Lorsch • Germany

PRISMAN GmbH
Otto-Hahn-Ring 6-18
D-64653 Lorsch • Germany
Tel.: +49 (0) 6251 866 980-0
Fax: +49 (0) 6251 866 980-77

info@prisman.de
www.prisman.de

Airflow Handy 2+, Airflow Handy 3.0 Perio Premium, Airflow One, Airflow Master (EMS Electro Medical Systems SA, Schweiz); Prophy-Mate neo⁴, Perio-Mate⁵ (NSK, Nakanishi Inc., Japan); Combi Touch, Starjet⁶

*(Mectron s.p.a., Italien); Air Max (Acteon Germany GmbH, Deutschland); PROPHY Flex 3 (KaVo Dental GmbH, Deutschland). Während der Tests wurde der Gebrauchsanweisung des Geräteherstellers Folge geleistet. Die oben genannten Pulver sind mit den getesteten Pulverstrahlgeräten kompatibel und freigegeben. Da die Bauart und Funktionsweise weiterer Geräte im Markt vergleichbar zu den getesteten ist, ist davon auszugehen, dass sich unsere Prophylaxepulver in allen Geräten problemlos anwenden lassen und sich daher die **Freigabe** über alle Geräte gleicher oder ähnlicher Bauart erstreckt.*

Generell sind die Gebrauchsanweisungen der Gerätehersteller sowie die Gebrauchsanweisungen der Prophylaxepulver zu beachten.

If you have further questions, please do not hesitate to contact us.

Bei Rückfragen stehen wir Ihnen gerne zur Verfügung.

Best regards

Mit freundlichen Grüßen

Lorsch, 26.04.2023


i.A. Dr. Kathrin Benzing
Quality Management

⁴ Es liegt eine Baugleichheit mit dem Gerät Smartgun One (Omnicent Dental-Handelsgesellschaft mbH, Germany) vor.

⁵ Es liegt eine Baugleichheit mit dem Gerät Smartgun Perio (Omnicent Dental-Handelsgesellschaft mbH, Germany) vor.

⁶ Es liegt eine Baugleichheit mit dem Gerät Proxeo Aura (W&H Dentalwerk Bürmoos GmbH, Österreich) vor.